

Notice of Allowability

Application No.

10/806,470

Examiner

Matthew F. DeSanto

Applicant(s)

AKAHOSHI, TAKAYUKI

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 7/20/06.
2. ☒ The allowed claim(s) is/are 1,3-11,13-26 and 28-38.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/08), Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☒ Interview Summary (PTO-413), Paper No./Mail Date 8/24/06.
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____.

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it **MUST** be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Adam Treiber on 8/24/06.

The application has been amended as follows:

Please amended Claim 35:

A method of phaco-emulsification, comprising: providing a needle having:

- a shaft;

- a tip, disposed at a distal end of said shaft, having:

- a distal terminus and at least two openings, defined by rounded edges,

- arranged near the terminus,

- an overall smooth, continuous curved shape, and

- a larger outer diameter than said shaft, and

- an aspiration lumen, extending through said shaft and said tip, in communication with the openings in the tip;

inserting said tip in an eye having a cataract;

vibrating said tip with substantially ultrasonic vibration;

breaking substantially said cataract into pieces with said tip; and

aspirating substantially said pieces through said aspiration lumen.

Reasons for Allowance

2. The following is an examiner's statement of reasons for allowance:

The subject matter of the independent claims could either not be found or was not suggested in the prior art of record. The subject matter not found in claims 1 and 11 are a phaco-emulsification needle with a shaft having a threaded portion and hub, a bolus metal tip with an opening at its distal terminus and wherein the edges of the opening are rounded and smooth. In claim 24 the subject matter not found was a phaco-emulsification needle with a shaft having a threaded portion and hub, a bolus metal tip with two opening at its distal terminus and wherein the edges of the openings are rounded and smooth. Claims 33-35 deal with the method of using a phaco-emulsification needle, wherein the needle is inserted into the eye and then vibrated by ultrasonic vibration to break up the cataract, wherein the pieces of the cataract are aspirated through the aspiration lumen of the needle.

The independent claims also include other patentable subject matter in combination with the other elements or steps of the claim not mention in the above paragraph.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Art Unit: 3763


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Matthew F. DeSanto whose telephone number is 571-272-4957. The examiner can normally be reached on Monday-Friday 9:30-6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick LUCCHESI can be reached on (571) 272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Matthew DeSanto
Art Unit 3763
August 30, 2006



Acting Supervisor
Du 3763